

Patient Functions


- How To: Enroll a Patient
- How to: Unenroll a Patient
- How To: Use Bypass
- How To: Suspend Tamper
- How To: Locate Patient


How To: Enroll a Patient






The enrollment window will pop up. Select the tag you wish to use from the drop-down list.



Enroll Patient
First Name
Middle Name
Last Name
Room
Not Set
Gender
Male
Flight Risk
Unit
Please select a unit
Tag
Please select a tag
Floor
Please select a tag
F1A26C
F136F6
B26E4F
B2032C
FUNCTIONS
© 2024 - McRoberts Technologies | build version: 1.0.0



Enter the patient's information. If you leave any required fields blank you will be prompted to fill them out. Only the "Middle Name" field is optional and can be left blank.
Enroll Patient
Enter Patient Information
Then Click Enroll

Patient will now have a status of "Active" and the total number of patients will be displayed on the dashboard. Other patient status are "Bypass" and "Tamper Suspended."

How to: Unenroll a Patient








After clicking "OK" you will receive a notification that the patient has been unenrolled. The page will then refresh and the patient will be removed from the patient directory. The tag status will change to "available" meaning it can be used on a new patient.

How To: Use Bypass


MyChild provides a ByPass that allows you to take a patient out of a protected area without setting off alarms. You must select a departing unit and a destination unit. These can be the same. The patient status changes from “Active” to “Bypass” when a patient is in ByPass.




Home
Patients
Tags
Floorplans
Reports
Admin
Help
Patient Directory
Enroll Patient
Low Battery
Granger, Hermione
UNIT: Mother / Baby
ROOM: 310
Jingle-Heimer-Schmidt, John
UNIT: Mother / Baby
ROOM: 311
LAST SEEN:
02/21/2024 11:07:21 AM
LOCATION:
Room 311
TAG: B21680
FUNCTIONS
TAG: E2484F
Rhodes, Daniel
BYPASS
SUSPEND TAMPER
UNENROLL
LOCATE
Dorman, Joe
UNIT: Mother / Baby
ROOM: 313
LAST SEEN:
02/21/2024 11:25:10 AM
LOCATION:
Room 313
TAG: E16269
FUNCTIONS
FLIGHT RISK
Low Battery
Doe, John
UNIT: Mother / Baby
ROOM: 315
LAST SEEN:
02/21/2024 11:05:41 AM
LOCATION:
Room 315
TAG: E10488
FUNCTIONS
Access all patient functions by clicking the pink FUNCTIONS button on the card.



Home
Patients
Tags
Floorplans
Reports
Admin
Help
Patient Directory
Enroll Patient
Low Battery
Granger, Hermione
UNIT: Mother / Baby
ROOM: 310
LAST SEEN:
02/21/2024 11:05:41 AM
LOCATION:
Room 310
TAG: B21580
FUNCTIONS
Jingle-Heimer-Schmidt, John
UNIT: Mother / Baby
ROOM: 311
LAST SEEN:
02/21/2024 11:07:21 AM
LOCATION:
Room 311
TAG: B21680
FUNCTIONS
TAG: E2484F
Rhodes, Daniel
BYPASS
SUSPEND TAMPER
UNENROLL
LOCATE
Click Bypass to continue
LOCATION:
Room 313
TAG: E16269
FUNCTIONS
FLIGHT RISK
Low Battery
Doe, John
UNIT: Mother / Baby
ROOM: 315
LAST SEEN:
02/21/2024 11:05:41 AM
LOCATION:
Room 315
TAG: E10488
FUNCTIONS

Enter a reason for the Bypass and a time limit.




Once the information is complete, click the "Put Patient In Bypass" button. This will change the patient status to Bypass. You will also see this change on the dashboard under the Patient Summary.

A level one notification will appear on the upper right corner of the screen telling you the Bypass has started once the patient leaves though the first bypass designed door.




This alarm indicates the patient has gone through the first door and the countdown has begun.


When the time allotted for the bypass is almost expired, a “BYPASS EXPIRING” alarm is triggered.



When the timer reaches zero, a “BYPASS EXPIRED” alarm is triggered. All bypass alarms (except BYPASS COMPLETE) allow you to edit the bypass. Changes can also be made on the patient card.



First, click the "Accept" button, then enter a message and click the green arrow to accept the alarm. The Bypass Expired alarm only happens if the time limit has run out and the patient has not passed through the second bypass designated door. If the patient arrives on time, you will receive a Bypass Complete alarm. This will show up in the upper right corner as a notification as it is a level one alarm. Click on the notification to view the alarm detail.




When you complete a bypass by clicking “Complete,” you are acknowledging that the patient has arrived at the destination unit. The patient status will return to "Active".

For additional help, call 1-800-776-7328 and press 2 for technical support.

Click here to download PDF.

How To: Suspend Tamper

Suspend Tamper Alarms: MyChild p
remove a cut-band or sensing tag tem
into SUSPEND TAMPER mode, the pat



When the time allotted for the suspen
EXPIRING" alarm is triggered. When t
is triggered. All suspend tamper alarn
be made on the patient card.

When you click Complete to complet
back on the patient.

Alarm Name	Level	Alarm Descriptio
SUSPEND TAMPER STARTED	1	This alarm notifie a patient has bee Suspend Tamper that the Suspend timer has started The patient statu from "Active" to Suspended."
SUSPEND	1	This alarm appea

Suspend Tamper PDF

When you click “Complete” to complete a Suspend Tamper, you are acknowledging that the tag is back on the patient.

How To: Locate Patient